

PHONAK
life is on


A Sonova brand

PerfectDry Lux.

User manual




x1



x1



x1



x1



EN


1

Connect the PerfectDry Lux. The blue indicator lights to show that it is powered. If no action occurs for 30 seconds, the indicator turns off.

2

Make sure the cover is closed when you start the cycle. Do not open the box when it is running, if you do so the PerfectDry Lux will stop in order to maintain safety.


The PerfectDry Lux is suitable for all types of hearing aids (including Rechargeable).



We recommend to clean hearing aids before dehumidification. Open the battery compartments (the batteries can be left inside the hearing aids during the cycle).

3

Press the touch button to start the cycle. The blue indicator flashes slowly. The PerfectDry Lux stops after 45 minutes.



Meaning of the light indicators

Blue Flashing > Running

Blue On > In standby mode

Function details

360° UV-C

The PerfectDry Lux has a short-wave UV-C lamp of which the 360° circular layout makes it possible to reach all the hearing aid surfaces. Hot air ventilation guarantees dehumidification of hearing aids. The lamp lights up for the first five minutes of the maintenance cycle, then turns off automatically.

Dehumidification

The PerfectDry Lux has a patented pulsed air dehumidification system. This means that the required power and drying time is reduced. The care chamber temperature is controlled by an electronic system in order to obtain a maximum temperature of 35 °C (95 °F) whatever the room temperature, thereby preventing damage to the hearing aids being cleaned.

In order to maintain optimum hearing aid quality, the daily use of the PerfectDry Lux is strongly recommended.

Safety instructions

Keep out of reach of children. Any work, other than cleaning and routine care by the client, must be carried out by the manufacturer's customer service or by an approved professional. Do not place the PerfectDry Lux, the power supply cable or the plug in water or any other liquid. If the power supply cable or the plug are damaged, do not use the device and replace them immediately. Replace your mains adapter with an original model. Place the PerfectDry Lux on a flat, dry and stable surface. Do not directly look at the UV-C lamp located in the dehumidification chamber. When disconnecting, pull the plug and not the cable. Never use detergent or corrosive products to clean your PerfectDry Lux. Failure to follow these instructions will void the guarantee. The manufacturer cannot be held liable for damage caused by incorrect use or handling.

Product care advice

Clean using a soft cloth. Do not clean the PerfectDry Lux using abrasive products. Do not spray liquids into the device. Keep the PerfectDry Lux closed when not in use to avoid impurities from being deposited.

Regulation applicable to Medical devices

This product meets the requirements of Regulation (EU) 2017/745 covering medical devices.

Guarantee

36 months - (cable and transformer included). In the event of a malfunction, contact your hearing-aid specialist. Proof of purchase will be requested for all work by the customer service during the guarantee period. This guarantee does not apply to defects caused by an accident or incident or similar damage, by introducing a liquid, negligence, unsuitable use, dismantling of any kind, lack of care or any other circumstance caused by the user. Furthermore, this guarantee will not apply to damage caused by lightning or any other electric voltage variation. As a precaution, please disconnect the PerfectDry Lux in the event of a thunderstorm.

Identification data

Meaning of abbreviations:

 This product is compliant with Regulation (EU) 2017/745 on medical devices.	
 Refer to the instruction manual/booklet	
 Manufacturer	 Date of manufacture
 Serial number	 Medical device
 Catalogue reference	5V-0.4A  Direct current
 Batch number	 Indoor use only
 The product is made in ABS with a recycling code of 7	 Sorting instructions
 DISPOSAL - Dispose of old appliances ecologically. Old devices contain recyclable materials which must be recovered. Therefore, make sure old devices are disposed of using the appropriate collection systems.	

Technical data

Functions

Internal (lxwxd): 70 x 21 x 51 mm
External (lxwxd): 93 x 48 x 76 mm
Weight (g): 144 g
UV lamp: 253.7 nm
Max. power: 4 W
Standby power: 0.04 W
Storage conditions from -20 °C to +65 °C

Drying temperature: 35 °C (95 °F)
Cycle duration: 45 min.
Ambient operating temperature:
From 12 °C to 40 °C
Power Supply:
Input: 100-240 Vac 50-60 Hz / Output: 5 Vdc
Connectors: USB cable - micro USB



I protect the environment, I disconnect my PerfectDry Lux once my hearing aids are clean!